

## Deutsche Akkreditierungsstelle GmbH

**Entrusted according to Section 8 subsection 1 AkkStelleG in connection with Section 1 subsection 1 AkkStelleGBV**

Signatory to the Multilateral Agreements of  
EA, ILAC and IAF for Mutual Recognition

# Accreditation



The Deutsche Akkreditierungsstelle GmbH attests

**Mahr GmbH**  
**Carl-Mahr-Straße 1, 37073 Göttingen**

that its calibration laboratory

**Mahr GmbH, Standort Esslingen**  
**Reutlinger Straße 48, 73728 Esslingen**

is competent under the terms of DIN EN ISO/IEC 17025:2005 to carry out calibrations in the following fields:

### Dimensional quantities

#### Length

- Gauge blocks
- Diameter
- Form error
- Length measuring instruments
- Length measuring devices
- Thread

The accreditation certificate shall only apply in connection with the notice of accreditation of 17.05.2017 with the accreditation number D-K-15074-02 and is valid until 16.05.2022. It comprises the cover sheet, the reverse side of the cover sheet and the following annex with a total of 4 pages.

Registration number of the certificate: **D-K-15074-02-00**

# Deutsche Akkreditierungsstelle GmbH

Office Berlin  
Spittelmarkt 10  
10117 Berlin

Office Frankfurt am Main  
Europa-Allee 52  
60327 Frankfurt am Main

Office Braunschweig  
Bundesallee 100  
38116 Braunschweig

The publication of extracts of the accreditation certificate is subject to the prior written approval by Deutsche Akkreditierungsstelle GmbH (DAkKS). Exempted is the unchanged form of separate disseminations of the cover sheet by the conformity assessment body mentioned overleaf.

No impression shall be made that the accreditation also extends to fields beyond the scope of accreditation attested by DAkKS.

The accreditation was granted pursuant to the Act on the Accreditation Body (AkkStelleG) of 31 July 2009 (Federal Law Gazette I p. 2625) and the Regulation (EC) No 765/2008 of the European Parliament and of the Council of 9 July 2008 setting out the requirements for accreditation and market surveillance relating to the marketing of products (Official Journal of the European Union L 218 of 9 July 2008, p. 30). DAkKS is a signatory to the Multilateral Agreements for Mutual Recognition of the European co-operation for Accreditation (EA), International Accreditation Forum (IAF) and International Laboratory Accreditation Cooperation (ILAC). The signatories to these agreements recognise each other's accreditations.

The up-to-date state of membership can be retrieved from the following websites:

EA: [www.european-accreditation.org](http://www.european-accreditation.org)

ILAC: [www.ilac.org](http://www.ilac.org)

IAF: [www.iaf.nu](http://www.iaf.nu)

# Deutsche Akkreditierungsstelle GmbH

## Annex to the Accreditation Certificate D-K-15074-02-00 according to DIN EN ISO/IEC 17025:2005

Period of validity: 17.05.2017 to 16.05.2022

Date of issue: 17.05.2017

Holder of certificate:

**Mahr GmbH**  
**Carl-Mahr-Straße 1, 37073 Göttingen**

with its calibration laboratory

**Mahr GmbH, Standort Esslingen**  
**Reutlinger Straße 48, 73728 Esslingen**

Head: Dipl.-Ing. Ulrich Börner  
Deputy: Martin Treiber  
Dipl.-Ing. (FH) Jürgen Bröckl

Accredited as calibration laboratory since: 24.03.1987

Calibration in the fields:

### **Dimensional quantities**

#### **Length**

- **Gauge blocks**
- **Diameter**
- **Form error**
- **Length measuring instruments**
- **Length measuring devices**
- **Thread**

Abbreviations used: see last page

Annex to the accreditation certificate D-K-15074-02-00

**Permanent Laboratory**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability <sup>1)</sup>	Remarks
<b>Length</b> Gauge blocks made of steel according to DIN EN ISO 3650:1999	0.5 mm to 100 mm featuring the nominal values of the standards	DAkkS-DKD-R 4-3 part 3.1:2010 Measurement of the deviation of the central length $l_c$ from the nominal value $l_n$ by comparison measurement	For the central length: $0.05 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$  For the deviations $f_0$ and $f_u$ from the central length: $0.05 \mu\text{m}$	$l$ = gauge block length Measuring surface quality as stated in Manuel resp. in the test specifications
	0.5 mm to 100 mm in the nominal dimensions, which differ of the standard (unusual nominal length)	Measurement of the deviations $f_0$ and $f_u$ from the central length by 5 points comparison For the smallest measurement uncertainties, the wringability and the wringing characteristics of both measuring surfaces must be checked using an appropriate optical flat	For the central length: $0.07 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$  For the deviations $f_0$ and $f_u$ from the central length: $0.05 \mu\text{m}$	
Gauge blocks made of ceramics or tungsten carbide according to DIN EN ISO 3650:1999	0.5 mm to 100 mm featuring the nominal values of the standards	Measurement of the deviation of the central length $l_c$ from the nominal value $l_n$ by comparison measurement	For the central length: $0.07 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot l$  For the deviations $f_0$ and $f_u$ from the central length: $0.05 \mu\text{m}$	
Pairs of gauge blocks made of steel or tungsten carbide according to DIN EN ISO 3650:1999	0.5 mm to 100 mm	DAkkS-DKD-R 4-3 part 3.1:2010 Measurement of the difference of the central lengths of gauge blocks with the same nominal length, respectively of difference of the central lengths up to $10 \mu\text{m}$ Measurement of the deviations $f_0$ and $f_u$ from the central length by 5 points comparison measurement	For the difference of the central length of the pairs: $0.03 \mu\text{m}$  For the deviations $f_0$ and $f_u$ from the central length: $0.03 \mu\text{m}$ (only for the nominal values $1.005 \text{ mm}$ and $1.01 \text{ mm}$ ) $0.05 \mu\text{m}$ (for the remaining values)	
Cylindrical setting gauges, Setting ring gauges Diameter	10 mm to 100 mm	DAkkS-DKD-R 4-3 part 4.1:2010	$0.4 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	$d$ = measured diameter
Straightness deviation and Parallelism deviation of surface lines			$0.5 \mu\text{m}$	
Roundness deviation			Cut-off wave number = 150 $0.1 \mu\text{m}$	
Cylindrical setting gauges, Setting plug gauges Diameter	3 mm to 100 mm	DAkkS-DKD-R 4-3 part 4.1:2010	$0.25 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	
Straightness deviation and Parallelism deviation of surface lines			$0.5 \mu\text{m}$	
Roundness deviation			Cut-off wave number = 150 $0.1 \mu\text{m}$	

<sup>1)</sup> The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of  $k=2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15074-02-00

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability <sup>1)</sup>	Remarks
Measuring pins Diameter	0.17 mm to 100 mm	DAkkS-DKD-R 4-3 part 4.2:2010	$0.25 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	$d$ = measured diameter
Straightness deviation and Parallelism deviation of surface lines	3 mm to 100 mm		0.5 $\mu\text{m}$	
Roundness deviation		Cut-off wave number = 150	0.1 $\mu\text{m}$	
Calipers for external, internal and depth dimensions	0 mm to 500 mm	DAkkS-DKD-R 4-3 part 9.1:2010	$15 \mu\text{m} + 15 \cdot 10^{-6} \cdot l$	$l$ = measured length
Micrometers	0 mm to 200 mm	DAkkS-DKD-R 4-3 part 10.1:2010	$2 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	scale interval 1 $\mu\text{m}$ , for higher scale intervals the meas- urement uncertainty will rise <hr/> 200 mm = final value of the measuring range
Reference gauges for micrometers	0 mm to 200 mm	DAkkS-DKD-R 4-3 part 4.4:2010	$1 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Dial gauges	to 30 mm	DAkkS-DKD-R 4-3 part 11.1:2010	$0.8 \mu\text{m} + 12 \cdot 10^{-6} \cdot l$	scale interval 1 $\mu\text{m}$ , for higher scale intervals the meas- urement uncertainty will rise
Dial indicators	to 3 mm	DAkkS-DKD-R 4-3 part 11.2:2010	0.5 $\mu\text{m}$	
Lever gauges	to 1.6 mm	DAkkS-DKD-R 4-3 part 11.3:2010	0.7 $\mu\text{m}$	
Inductive probe and measuring device	to 10 mm	DAkkS-DKD-R 4-3 part 14.1:2010	0.5 $\mu\text{m}$	
Inductive probe without measuring device	to 10 mm		1.2 $\mu\text{m}$	
Inkrementale Messtaster	0 mm to 100 mm	According to VDI/VDE/DGQ 2618 part 14.1:2010	$0.35 \mu\text{m} + 12 \cdot 10^{-6} \cdot l$	
Height gauges / Heigh measuring machines	0 mm to 1000 mm	VDI/VDE/DGQ 2618 part 16.1:2009	$1.7 \mu\text{m} + 1.2 \cdot 10^{-6} \cdot l$	

<sup>1)</sup> The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of  $k=2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

**Annex to the accreditation certificate D-K-15074-02-00**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability <sup>1)</sup>	Remarks
Thread gauges (single-start cylindrical external and internal threads with straight flanks and symmetrical profile)				
External thread Simple pitch diameter with nominal lead 0.25 mm to 6 mm	Nominal diameter 3 mm to 100 mm	DAkkS-DKD-R 4-3 part 4.8:2010, Option 1	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$	$d = \text{pitch diameter}$
Internal thread Simple pitch diameter with nominal lead 0.25 mm to 6 mm	Nominal diameter 5 mm to 100 mm	DAkkS-DKD-R 4-3 part 4.9:2010, Option 1	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$	

**Abbreviations used:**

DAkkS-DKD-R Guideline on Deutsche Akkreditierungsstelle GmbH

<sup>1)</sup> The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of  $k=2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.