

## Deutsche Akkreditierungsstelle GmbH

**Entrusted according to Section 8 subsection 1 AkkStelleG in connection with Section 1 subsection 1 AkkStelleGBV**

Signatory to the Multilateral Agreements of EA, ILAC and IAF for Mutual Recognition

# Accreditation



The Deutsche Akkreditierungsstelle GmbH attests that the calibration laboratory

**Mahr GmbH**  
**Carl-Mahr-Straße 1, 37073 Göttingen**

is competent under the terms of DIN EN ISO/IEC 17025:2018 to carry out calibrations in the following fields:

### Dimensional quantities

#### Length

- Roughness
- Form error
- Contours
- Stylus instruments <sup>a)</sup>
- Length measuring devices <sup>a)</sup>

<sup>a)</sup> also on-site calibration

The accreditation certificate shall only apply in connection with the notice of accreditation of 22.07.2021 with the accreditation number D-K-15074-01. It comprises the cover sheet, the reverse side of the cover sheet and the following annex with a total of 6 pages.

Registration number of the certificate: **D-K-15074-01-00**

Braunschweig,  
22.07.2021

Dr Heike Manke  
Head of Division

Translation issued:  
22.07.2021



Head of Division

*The certificate together with the annex reflects the status as indicated by the date of issue.  
The current status of any given scope of accreditation may be found respectively in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH <https://www.dakks.de/en/content/accredited-bodies-dakks>.*

This document is a translation. The definitive version is the original German accreditation certificate.  
See notes overleaf.

## Deutsche Akkreditierungsstelle GmbH

### Annex to the Accreditation Certificate D-K-15074-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: **22.07.2021**

Date of issue 22.07.2021

Holder of certificate:

**Mahr GmbH**  
**Carl-Mahr-Straße 1, 37073 Göttingen**

Calibration in the fields:

#### Dimensional quantities

##### Length

- Roughness
- Form error
- Contours
- Stylus instruments <sup>a)</sup>
- Length measuring devices <sup>a)</sup>

<sup>a)</sup> also on-site calibration

*The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories. Laboratories that conform to the requirements of this standard, operate generally in accordance with the principles of DIN EN ISO 9001.*

*The certificate together with the annex reflects the status as indicated by the date of issue.*

*The current status of any given scope of accreditation may be found respectively in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH <https://www.dakks.de/en/content/accredited-bodies-dakks>.*

Abbreviations used: see last page

**Page 1 of 6**

Annex to the accreditation certificate D-K-15074-01-00

**Permanent Laboratory**

Measurement quantity / Calibration item	Range	Calibration and Measurement Capabilities (CMC)		Remarks
		Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	
<b>Length</b>				
Groove depth $P_t$ and $d$ on depth setting standards	0.2 µm to (7.9 µin)      0.8 µm to (31.5 µin) 0.8 µm to (31.5 µin)      1.5 µm to (59.1 µin) 1.5 µm to (59.1 µin)      3.5 µm to (137.8 µin) 3.5 µm to (137.8 µin)      12 µm to (472.5 µin)	DIN EN ISO 3274:1998 DIN EN ISO 4287:2010 DIN EN ISO 5436-1:2000	0.03 µm (1.2 µin) 0.04 µm (1.6 µin) 0.04 µm (1.6 µin) 0.05 µm (2.0 µin)	
Roughness on extra fine roughness standards		DIN 4768:1990 DIN EN ISO 3274:1998		
$R_a$		DIN EN ISO 4287:2010		
$R_z$	0.025 µm to (1.0 µin)      0.1 µm to (3.9 µin)	DIN EN ISO 4288:1998 DIN EN ISO 16610-21:2013	0.08 · $R_a$	
$R_{max}$	0.15 µm to (5.9 µin)      0.15 µm to (5.9 µin)      0.8 µm to (31.5 µin)		0.09 · $R_z$ 0.10 · $R_{max}$	
Roughness on roughness standards		DIN 4768:1990 DIN EN ISO 3274:1998		
$R_a$	0.1 µm to (3.9 µin)      4 µm to (157.5 µin)	DIN EN ISO 4287:2010 DIN EN ISO 4288:1998	0.05 · $R_a$	
$R_z$	0.8 µm to (31.5 µin)      20 µm to (787.5 µin)	DIN EN ISO 16610-21:2013	0.05 · $R_z$	
$R_{max}$	0.8 µm to (31.5 µin)      20 µm to (787.5 µin)		0.05 · $R_{max}$	
Roughness on roughness standards	On surfaces in the range	DIN 4768:1990 DIN EN ISO 13565-1:1998 DIN EN ISO 13565-2:1998		
$R_{pk}$	0.1 µm ≤ $R_a$ ≤ (3.9 µin)      4 µm to (157.5 µin)		0.04 · $R_z$	Relative measuring uncertainty relative to $R_z$
$R_k$			0.05 · $R_z$	
$R_{vk}$			0.04 · $R_z$	
$M_{rl}$	0.8 µm ≤ $R_z$ ≤ (31.5 µin)      20 µm to (787.5 µin)		4 %	Absolute measuring uncertainty relative to 100% material ratio
$M_{r2}$			6 %	
Roughness on roughness standards		DIN 4768:1990 DIN EN ISO 3274:1998		If necessary, the filter cutoff wavelength $\lambda_c$ can be used one level lower or higher than as per ISO 4288:1998
$R_a$	0.1 µm to (3.9 µin)      4 µm to (157.5 µin)	DIN EN ISO 4287:2010 DIN EN ISO 4288:1998	0.03 · $R_a$	
$R_z$	0.8 µm to (31.5 µin)      20 µm to (787.5 µin)	DIN EN ISO 16610-21:2013	0.03 · $R_z$	
$R_{max}$	0.8 µm to (31.5 µin)      20 µm to (787.5 µin)		0.03 · $R_{max}$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Date of issue: 22.07.2021

Valid from: 22.07.2021

Page 2 of 6

Annex to the accreditation certificate D-K-15074-01-00

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)					
Measurement quantity / Calibration item	Range		Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Stylus instruments (Surface) to DIN EN ISO 3274:1998 <i>Pt</i> and <i>d</i>	0.2 µm to 12 µm (7.9 µin) (472.5 µin)		DKD-R 4-2 part 2:2018 DIN EN ISO 12179:2000	$U_{\text{standard}} + 0.01 \mu\text{m}$ ( $U_{\text{standard}} + 0.4 \mu\text{in}$ )	$U_{\text{standard}}$ is the measuring uncertainty of the standards used.
<i>Ra</i>	0.1 µm to 4 µm (3.9 µin) (157.5 µin)			$U_{\text{standard}} + 0.01 \cdot Ra$	Smaller measuring ranges for which standards are available can also be calibrated.
<i>Rz</i>	0.8 µm to 20 µm (31.5 µin) (787.5 µin)			$U_{\text{standard}} + 0.01 \cdot Rz$	
<i>Rmax</i>	0.8 µm to 20 µm (31.5 µin) (787.5 µin)			$U_{\text{standard}} + 0.01 \cdot Rmax$ ( $U_{\text{standard}} + 0.4 \mu\text{in} \cdot Rmax$ )	
Stylus instruments (Contour) to DIN EN ISO 3274 and VDI/VDE/DGQ 2604 Bl.1			MK03/07:2021		The measuring uncertainty of mechanical scanning of contour standards and contour stylus instruments of the Mahr GmbH
Distance X	to 100 mm (3.94 in)			1.25 µm (49.2 µin)	
Distance Z	to 10 mm (0.39 in)			1.1 µm (43.3 µin)	
Radii	2 mm to 12 mm (0.079 in) (0.47 in)			1.6 µm (63.0 µin)	
Angle	40° to 135°			0.025°	
Straightness	to 20 µm (787.5 µin)			0.36 µm (14.2 µin)	
Roundness standards Roundness deviation	to 0.1 µm (3.9 µin)		DIN ISO 1101:2014	0.025 µm (1.0 µin)	Diameter: 3 mm to 100 mm (0.12 to 3.94 in)
Magnification standards Roundness deviation for cylinder with flat area (flick)	0.5 µm to 20 µm (19.7 µin) (787.5 µin)			0.05 µm + $2.5 \cdot 10^{-2} \cdot RON_t$ (2.0 µin + $2.5 \cdot 10^{-2} \cdot RON_t$ )	Diameter: 3 mm to 100 mm (0.12 to 3.94 in)
Magnification standards Roundness deviation Multi-wave standard	to 20 µm (787.5 µin)			0.1 µm + $2.5 \cdot 10^{-2} \cdot RON_t$ (3.9 µin + $2.5 \cdot 10^{-2} \cdot RON_t$ )	Diameter: 50 mm to 150 mm (1.97 to 5.91 in)
Cylinder square Roundness deviation	to 20 µm (787.5 µin)		DIN ISO 1101:2014	0.1 µm + $2.5 \cdot 10^{-2} \cdot RON_t$ (3.9 µin + $2.5 \cdot 10^{-2} \cdot RON_t$ )	Diameter: 3 mm to 100 mm (0.12 to 3.94 in)
Straightness deviation of the generatrices	to 20 µm (787.5 µin)			0.2 µm + $2.0 \cdot 10^{-2} \cdot STR_t$ (7.9 µin + $2.0 \cdot 10^{-2} \cdot STR_t$ )	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Date of issue: 22.07.2021

Valid from: 22.07.2021

Page 3 of 6

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Parallelism deviation of the generatrices	to 20 µm (787.5 µin)		0.3 µm + 1.5 · 10 <sup>-2</sup> · PARt (11.8 µin + 1.5 · 10 <sup>-2</sup> · PARt)	Length: 10 mm to 400 mm (0.39 to 15.75 in)  RONt = roundness deviation STRt = Straightness deviation PARt = Parallelism deviation CYLt = Cylindricity deviation
Cylindricity deviation	to 20 µm (787.5 µin)		0.4 µm + 3.0 · 10 <sup>-2</sup> · CYLt (15.8 µin + 3.0 · 10 <sup>-2</sup> · CYLt)	
Contour standards		Substitution measurement with reference contour standard		
X length Lateral distances	5 mm to 100 mm (0.20 in) to 100 mm (3.94 in)	Procedure according to DIN ISO/TS 15530-3:2008	0.6 µm (23.6 µin)	
Z length Vertical distances	to 10 mm (0.39 in)		0.75 µm (29.5 µin)	
Radii	2 mm to 12 mm (0.079 in) to 12 mm (0.47 in)		0.75 µm (29.5 µin)	
Angles	40° to 135°		0.01°	
Dial gauge checkers	to 100 mm (3.94 in)	MK03/05:2014 Calibration with traceable electronic linear reference gauge	0.22 µm (8.7 µin)	
Horizontal Length measuring machines	0 mm to 1000 mm (0 µin) to 1000 mm (39.37 in)	VDI/VDE/DGQ 2618 part 17.1:2014	0.08 µm + 0.7 · 10 <sup>-6</sup> · l (3.1 µin + 0.7 · 10 <sup>-6</sup> · l)	<i>l</i> = measured length The measurement uncertainty of the length measurement uncertainty in mechanical probing of gauge blocks and is valid for horizontal length measuring machines of the Mahr GmbH
	> 1000 mm to 2000 mm (> 39.37 in) to 2000 mm (78.74 in)		0.1 µm + 0.5 · 10 <sup>-6</sup> · l (3.9 µin + 0.5 · 10 <sup>-6</sup> · l)	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of *k* = 2 unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

**On-site Calibration**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>Length</b>				
Stylus instruments (surface) to DIN EN ISO 3274:1998	0.2 µm to 12 µm (7.9 µin) (472.5 µin)	DKD-R 4-2 part 2:2018 DIN EN ISO 12179:2000	$U_{\text{standard}} + 0.01 \mu\text{m}$ ( $U_{\text{standard}} + 0.4 \mu\text{in}$ )	$U_{\text{standard}}$ is the measuring uncertainty of the standards used.
$P_t$	0.1 µm to 4 µm (3.9 µin) (157.5 µin)		$U_{\text{standard}} + 0.01 \cdot R_a$ ( $U_{\text{standard}} + 0.4 \mu\text{in} \cdot R_a$ )	Smaller measuring ranges for which standards are available can also be calibrated.
$R_a$	0.8 µm to 20 µm (31.5 µin) (787.5 µin)		$U_{\text{standard}} + 0.01 \cdot R_z$ ( $U_{\text{standard}} + 0.4 \mu\text{in} \cdot R_z$ )	
$R_z$	0.8 µm to 20 µm (31.5 µin) (787.5 µin)		$U_{\text{standard}} + 0.01 \cdot R_{max}$ ( $U_{\text{standard}} + 0.4 \mu\text{in} \cdot R_{max}$ )	
$R_{max}$				
Stylus instruments (Contour) to DIN EN ISO 3274 and VDI/VDE/DGQ 2604 Bl.1		MK03/07:2021		The measuring uncertainty of mechanical scanning of contour standards and contour stylus instruments of the Mahr GmbH
Distance X	to 100 mm (3.94 in)		1.25 µm (49.2 µin)	
Distance Z	to 10 mm (0.39 in)		1.1 µm (43.3 µin)	
Radii	2 mm to 12 mm (0.079 in) (0.47 in)		1.6 µm (63.0 µin)	
Angle	40° to 135°		0.025°	
Straightness	to 20 µm (787.5 µin)		0.36 µm (14.2 µin)	
Dial gauge checkers	to 100 mm (3.94 in)	MK03/05:2014 Calibration with traceable electronic linear reference gauge	0.22 µm (8.7 µin)	
Horizontal Length measuring machines	0 mm to 1000 mm (0 µin) (39.37 in)	VDI/VDE/DGQ 2618 part 17.1:2014	0.08 µm + 0.7 · 10 <sup>-6</sup> · l (3.1 µin + 0.7 · 10 <sup>-6</sup> · l)	$l$ = measured length The measurement uncertainty of the length measurement uncertainty in mechanical probing of gauge blocks and is valid for horizontal length measuring machines of the Mahr GmbH
	> 1000 mm to 2000 mm (> 39.37 in) (78.74 in)		0.1 µm + 0.5 · 10 <sup>-6</sup> · l (3.9 µin + 0.5 · 10 <sup>-6</sup> · l)	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.



**Annex to the accreditation certificate D-K-15074-01-00**

**Abbreviations used:**

CMC	Calibration and measurement capabilities
DIN	Deutsches Institut für Normung e.V.
DKD-R	Guideline on Deutscher Kalibrierdienst (DKD), published by Physikalisch-Technischen Bundesanstalt
VDE	Verband der Elektrotechnik, Elektronik und Informationstechnik
VDI	Verein Deutscher Ingenieure
MK	Calibration instruction of the Mahr GmbH

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Date of issue: 22.07.2021  
Valid from: 22.07.2021

**Page 6 of 6**